

## CERTIFICATE OF CALIBRATION / VALIDATION

Customer	HighQ Pharmaceuticals	Certificate #	MVLT_COC_090125_17628
Address	Plot No. 224 & 225/1, Sector 23 Korangi	Date of Issue	09-01-25
E mail	<a href="mailto:purchase@highqpharma.com">purchase@highqpharma.com</a>	Date of Expiry	08-01-26

Unit Under Test UUT			
User ID	STR-THM-19	Sub Location	Sterile Filling Area
Location	Sterile Cephalosporin	Instrument Type	Hygrometer
Model	HTC-1	Make	China
Range	-10~50°C & 0-100 %RH	Accuracy/Resolution	0.1°C % 1 %RH
Serial #	N/A	Condition	Used
Least Count	0.1°C & 1 %RH	Tolerance Limit	± 1°C & 5 %RH

Master / Testing / Reference Equipment Used			
Hygrometer			
Model	HUATO	Serial #	100FA188
Range	-20 To 70 C 0 to 100 %RH	Accuracy	0.1 C / 0.1 %RH
Traceability Ref #	STL/HYG/2024/03	Cal Due date	2025-11-21
Humidity Controlled Box			
Model	N/A	Serial #	N/A
Range	N/A	Accuracy	N/A
Traceability Ref #	N/A	Cal Due date	N/A

Test Performed / Readings			
Master Reading	UUT Readings	Difference	Review
°C	°C	°C	°C
21.0	21.0	0.00	Pass
%RH	%RH	%RH	%RH
55	55	0.00	Pass

\*UUT = Unit under Test \* N/A= Not applicable

**Remarks:** Above given Hygrometer working as per manufacture specifications.



Noman Khan (Test Eng.)

PNAC: ISO 17025 Certified Assessor  
ISO/IEC17025/APLACTR001

Company Stamp

HighQ Representative