

CERTIFICATE OF CALIBRATION / VALIDATION

Customer	TheSearleCompany Ltd	Certificate #	MVLT_COC_191122_1167
Address	F-319, S.I.T.E Karachi	Date of Issue	19-11-22
E mail	Mustafain.baig@searlecompany.com	Date of Expiry	18-05-23

Unit Under Test UUT			
User ID	CINC-ML-001	Sub Location	Instrument Room
Location	Micro Lab	Instrument Type	Cool Incubator
Model	MIR-253	Make	Sanyo Japan
Range	10~50 °C	Accuracy/Resolution	1 °C
Serial #	N/A	Condition	PASS
Least Count	1 °C	Tolerance Limit	± 1 °C

Master / Testing / Reference Equipment Used			
Temperature Data logger			
Model	GL-800	Serial #	70410776
Range	0-1200 °C	Accuracy	0.1 °C
Traceability Ref #	ILS/M/0920/250/04	Cal Due date	2023-09-27

Test Performed / Readings			
Master Reading	UUT Readings	Difference	Review
°C	°C	°C	°C
22.0	22.0	0.00	Pass
22.0	22.0	0.00	Pass
22.0	22.0	0.00	Pass

*UUT = Unit under Test * N/A= Not applicable

Remarks: Above given Cool Incubator working as per manufacture specifications.



Noman Khan (Test Eng.)

*PNAC: ISO 17025 Certified Assessor
ISO/IEC17025/APLACTR001*

Company Stamp

TheSearleCompany Representative