

CERTIFICATE OF CALIBRATION / VALIDATION

Customer	TheSearleCompany Ltd	Certificate #	MVLT_COC_191122_1143
Address	F-319, S.I.T.E Karachi	Date of Issue	19-11-22
E mail	Mustafain.baig@searlecompany.com	Date of Expiry	18-05-23

Unit Under Test UUT			
User ID	CG-ML-001	Sub Location	Media Room
Location	Micro Lab	Instrument Type	Compound Gauge
Model	OXID	Make	OXID,OXOID
Range	-1~1 Bar	Accuracy/Resolution	0.2 Bar
Serial #	N/A	Condition	PASS
Least Count	0.2 Bar	Tolerance Limit	± 0.2 Bar

Master / Testing / Reference Equipment Used			
Digital Pressure Gauge			
Model	ADDITEL 680	Serial #	190109
Range	0-42000 Bar	Accuracy	0.05 %
Traceability Ref #	ILS/0422/195/01	Cal Due date	2024-03-04
Hand Pressure Pump			
Model	PV-212	Serial #	N/A
Range	0-700 Bar	Accuracy	N/A Bar
Traceability Ref #	N/A	Cal Due date	N/A

Test Performed / Readings			
Master Reading	UUT Readings	Difference	Review
Bar	Bar	Bar	Bar
0.2	0.2	0.00	Pass
0.6	0.6	0.00	Pass
1.0	1.0	0.00	Pass

*UUT = Unit under Test * N/A= Not applicable

Remarks: Above given Compound Gauge working as per manufacture specifications.



Noman Khan (Test Eng.)

PNAC: ISO 17025 Certified Assessor
ISO/IEC17025/APLACTR001

Company Stamp

TheSearleCompany Representative