

CERTIFICATE OF CALIBRATION / VALIDATION

Customer	TheSearleCompany Ltd	Certificate #	MVLT_COC_211122_1446
Address	F-319, S.I.T.E Karachi	Date of Issue	18-10-22
E mail	Mustafain.baig@searlecompany.com	Date of Expiry	17-04-23

Unit Under Test UUT			
User ID	CM-SU-001	Sub Location	WFI PLANT
Location	STERILE UTILITY	Instrument Type	Conductivity Meter
Model	HANNA	Make	HANNA
Range	MULTIPLE FS	Accuracy/Resolution	1 FS
Serial #	N/A	Condition	PASS
Least Count	1 FS	Tolerance Limit	± 1 FS

Master / Testing / Reference Equipment Used			
Conductivity Solution			
Model	Extech	Serial #	000002
Range	1280 , 1413 µs/cm	Accuracy	1 %
Traceability Ref #	EXTECH	Cal Due date	2022-12-30

Test Performed / Readings			
Master Reading	UUT Readings	Difference	Review
µs	µs	µs	µs
0	0	0.00	PASS
1413	1413	0.00	PASS
1413	1413	0.00	PASS
1413	1413	0.00	PASS

*UUT = Unit under Test * N/A= Not applicable

Remarks: Above given Conductivity Meter working as per manufacture specifications.



Noman Khan (Test Eng.)

*PNAC: ISO 17025 Certified Assessor
ISO/IEC17025/APLACTR001*

Company Stamp

TheSearleCompany Representative