

## CERTIFICATE OF CALIBRATION / VALIDATION

Customer	TheSearleCompany Ltd	Certificate #	MVLT_COC_211122_1461
Address	F-319, S.I.T.E Karachi	Date of Issue	17-11-22
E mail	<a href="mailto:Mustafain.baig@searlecompany.com">Mustafain.baig@searlecompany.com</a>	Date of Expiry	16-05-23

Unit Under Test UUT			
User ID	CG-PL-011	Sub Location	AUTOCLAVE AREA
Location	PEDITRAL LIQUID	Instrument Type	Compound Gauge
Model	BLD	Make	BLD
Range	-0.1 ~ 0.9 Mpa	Accuracy/Resolution	0.01 Mpa
Serial #	N/A	Condition	PASS
Least Count	0.01 Mpa	Tolerance Limit	± 0.01 Mpa

Master / Testing / Reference Equipment Used			
Digital Pressure Gauge			
Model	ADDITEL 680	Serial #	190109
Range	0-42000 Bar	Accuracy	0.05 %
Traceability Ref #	ILS/0422/195/01	Cal Due date	2024-03-04
Humidity Controlled Box			
Model	N/A	Serial #	N/A
Range	N/A	Accuracy	N/A
Traceability Ref #	N/A	Cal Due date	N/A

Test Performed / Readings			
Master Reading	UUT Readings	Difference	Review
Mpa	Mpa	Mpa	Mpa
0.0	0.0	0.00	PASS
0.2	0.2	0.00	PASS
0.4	0.4	0.00	PASS
0.6	0.6	0.00	PASS

\*UUT = Unit under Test \* N/A= Not applicable

**Remarks:** Above given Compound Gauge working as per manufacture specifications.



\_\_\_\_\_  
Noman Khan (Test Eng.)

*PNAC: ISO 17025 Certified Assessor  
ISO/IEC17025/APLACTR001*

\_\_\_\_\_  
Company Stamp

\_\_\_\_\_  
TheSearleCompany Representative