

CERTIFICATE OF CALIBRATION / VALIDATION

| | | | |
|----------|--|----------------|----------------------|
| Customer | TheSearleCompany Ltd | Certificate # | MVLT_COC_211122_1449 |
| Address | F-319, S.I.T.E Karachi | Date of Issue | 18-10-22 |
| E mail | Mustafain.baig@searlecompany.com | Date of Expiry | 17-04-23 |

| Unit Under Test UUT | | | |
|---------------------|-----------------|---------------------|---------------------|
| User ID | CM-SU-004 | Sub Location | D.I TANK |
| Location | STERILE UTILITY | Instrument Type | Conductivity Meter |
| Model | ROC | Make | ROC |
| Range | MULTIPLE | Accuracy/Resolution | 1 FS |
| Serial # | N/A | Condition | PASS |
| Least Count | 1 FS | Tolerance Limit | $\pm 1 \mu\text{s}$ |

| Master / Testing / Reference Equipment Used | | | |
|---|------------------------------|--------------|------------|
| Conductivity Solution | | | |
| Model | Extech | Serial # | 000002 |
| Range | 1280 , 1413 $\mu\text{s/cm}$ | Accuracy | 1 % |
| Traceability Ref # | EXTECH | Cal Due date | 2022-12-30 |

| Test Performed / Readings | | | |
|---------------------------|--------------|------------|--------|
| Master Reading | UUT Readings | Difference | Review |
| 1413 | 1413 | 0.00 | PASS |
| 1413 | 1413 | 0.00 | PASS |
| 1413 | 1413 | 0.00 | PASS |
| 1413 | 1413 | 0.00 | PASS |

*UUT = Unit under Test * N/A= Not applicable

Remarks: Above given Conductivity Meter working as per manufacture specifications.



Noman Khan (Test Eng.)

*PNAC: ISO 17025 Certified Assessor
ISO/IEC17025/APLACTR001*

Company Stamp

TheSearleCompany Representative