

## CERTIFICATE OF CALIBRATION / VALIDATION

Customer	TheSearleCompany Ltd	Certificate #	MVLT_COC_131222_1758
Address	F-319, S.I.T.E Karachi	Date of Issue	30-10-22
E mail	<a href="mailto:Mustafain.baig@searlecompany.com">Mustafain.baig@searlecompany.com</a>	Date of Expiry	29-03-23

Unit Under Test UUT			
User ID	CM-CU-006	Sub Location	Bioler Area
Location	CENTRAL UTILITY	Instrument Type	TDs Meter
Model	DIST 2	Make	HANNA
Range	0-1999 PPM	Accuracy/Resolution	1 PPM
Serial #	N/A	Condition	PASS
Least Count	1 PPM	Tolerance Limit	± 1 PPM

Master / Testing / Reference Equipment Used			
Conductivity Solution			
Model	Extech	Serial #	000002
Range	1280 , 1413 µs/cm	Accuracy	1 %
Traceability Ref #	EXTECH	Cal Due date	2022-12-30

Test Performed / Readings			
Master Reading	UUT Readings	Difference	Review
PPM	PPM	PPM	PPM
1413	1412	1.00	Pass
1280	1280	0.00	Pass

\*UUT = Unit under Test \* N/A= Not applicable

**Remarks:** Above given TDs Meter working as per manufacture specifications.



\_\_\_\_\_  
Noman Khan (Test Eng.)

*PNAC: ISO 17025 Certified Assessor  
ISO/IEC17025/APLACTR001*

\_\_\_\_\_  
Company Stamp

\_\_\_\_\_  
TheSearleCompany Representative