

CERTIFICATE OF CALIBRATION / VALIDATION

Customer	HighQ Pharmaceuticals	Certificate #	MVLT_COC_090125_17620
Address	Plot No. 224 & 225/1, Sector 23 Korangi	Date of Issue	09-01-25
E mail	purchase@highqpharma.com	Date of Expiry	08-01-26

Unit Under Test UUT			
User ID	ENG-PG-157	Sub Location	High-Q Steam Generator
Location	Engineering	Instrument Type	Pressure Gauge
Model	WISE	Make	WISE
Range	0-2 Mpa	Accuracy/Resolution	0.05 Mpa
Serial #	N/A	Condition	Used
Least Count	0.05 Mpa	Tolerance Limit	± 0.05 Mpa

Master / Testing / Reference Equipment Used			
Digital Pressure Gauge			
Model	ADDITEL 680	Serial #	190109
Range	0-4200 Bar	Accuracy	0.05 %
Traceability Ref #	STL/2024/190109	Cal Due date	2025-02-02
Hand Pressure Pump			
Model	PV-212	Serial #	N/A
Range	0-700 Bar	Accuracy	N/A Bar
Traceability Ref #	N/A	Cal Due date	N/A

Test Performed / Readings			
Master Reading	UUT Readings	Difference	Review
Mpa	Mpa	Mpa	Mpa
0.5	0.5	0.00	Pass
1.0	1.0	0.00	Pass
1.5	1.5	0.00	Pass

*UUT = Unit under Test * N/A= Not applicable

Remarks: Above given Pressure Gauge working as per manufacture specifications.



Noman Khan (Test Eng.)

*PNAC: ISO 17025 Certified Assessor
ISO/IEC17025/APLACTR001*

Company Stamp

HighQ Representative